

J. Closed Beam X-ray Machine

LAB/ROOM: INSPECTED BY: DATE:

Item	Subject	Yes, No, N.A.	Comment
J1	Radiation hazard warning sign is posted at the entrance.		
J2	The room is accessible to authorized users only.		
J3	Copy of the valid Irradiating Apparatus Licence is displayed in the room.		
J4	Standard operating procedures / guidelines or safety rules displayed in the room.		
J5	Chamber enclosing the X-ray tube housing, sample and detector are intact.		
J6	Interlock of the chamber (to prevent X-ray exposure when the chamber doors are opened) functions.		
J7	There is visible signal on the machine to indicate the X-ray (tube) is on and this signal light functions.		
J8	There is visible signal on the machine to indicate the shutter is open and this signal light functions.		
J9	Other		

Endorsed By _____ Date _____
 Signature & Name