

F. High Power Laser (Class 3b and/or 4)

LAB/ROOM: INSPECTED BY: DATE:

Item	Subject	Yes, No, N.A.	Comment
F1	Proper laser hazard warning sign is posted at the entrance.		
F2	Laser warning light is installed at the entrance and functions.		
F3	The room is accessible to authorized users only.		
F4	The laser is affixed with laser warning label(s)		
F5	Standard operating procedures / guidelines or safety rules displayed in the room.		
F6	The laser is situated or mounted in its original position.		
F7	The laser beam is directed away from users or other occupiers of the room.		
F8	No flammable substances are placed close to the laser.		
F9	The laser beam is not directing at the human eye level.		
F10	The key for switching on/off the laser is properly kept when not in use.		
F11	The protective housing or enclosure of the laser is intact.		
F12	Interlock of the protective housing or enclosure functions.		
F13	Interlock of the room door functions.		
F14	There are sufficient beam stops, screens or barriers available for use.		
F15	Reflective surfaces in the vicinity not related to the work removed or covered.		
F16	There is sufficient lighting in the room.		
F17	There is sufficient local exhaust ventilation associated with the laser operation.		
F18	An appropriate fire extinguisher is available in the room containing a class 4 laser.		
F19	Suitable laser eye protectors are available for use.		
F20	Laser eye protectors are kept in good conditions and without observable defects.		
F21	Other		

Endorsed By _____ Date _____

Signature & Name