

D. Compressed gas safety

LAB/ROOM: INSPECTED BY: DATE:

Item	Subject	Yes, No, N.A.	Comment
D1	Compressed gas cylinders are properly secured		
D2	The number of compressed gas cylinders is within the exemption quantity		
D3	Gas cylinders are located away from heat sources such as direct exposure to sunlight, ovens etc		
D4	Oxidising and flammable gases are located away from electric sockets and switches		
D5	Any local exhaust system for toxic gas is functioning properly		
D6	Gas cylinder trucks are available for transporting cylinders		Remark: "available" should read "used".
D7	Flashback arrestors are installed for gas cylinders used to produce naked flames		Remark: For combustible gas only.
D8	Proper keys are available to open the main valves of has cylinders		Remark: "has" should read "gas".
D9	Cryogenic gloves and aprons, if required, are available for handling cryogenic materials		
D10	Privately owned gas cylinders are pressure-tested every 5 years		
D11	Other		

Endorsed By _____ Date _____
 Signature & Name